

BLANK PAGE



IS: 4162 (Part 3) - 1985

Indian Standard

REAFFIRMED

<u>...</u> 2007

SPECIFICATION FOR GRADUATED PIPETTES

PART 3 PIPETTES FOR WHICH A WAITING TIME OF 15s IS SPECIFIED

(First Revision)

First Reprint FEBRUARY 1996

UDC 542.395 [666.172.7]:531.732

© Copyright 1986

BUREAU OF INDIAN STANDARDS MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG NEW DELHI 110002

AMENDMENT NO. 1 JULY 2007 TO

IS 4162 (PART 3): 1985 SPECIFICATION FOR GRADUATED PIPETTES

PART 3 PIPETTES FOR WHICH A WAITING TIME OF 15s IS SPECIFIED

(First Revision)

(Page 5, clause 7.1.1) — Substitute the following for the existing clause:

'7.1.1 Graduated pipettes shall be constructed from glass of suitable chemical and thermal properties, shall be as far as possible, free from visible defects and shall be reasonably free from internal stress. When graded according to the method prescribed in IS 2303 (Part 1/Sec 1): 1994*, the pipettes shall conform to Class HGB 1 of the glass.'

(Page 5, footnote marked *) — Substitute the following for the existing:

**Grading glass for alkalinity: Part 1 Hydrolytic resistance, Section 1 Hydrolytic resistance of glass grains at 98°C. Method of test and classification (first revision).*

(Page 8, clause 13.1.1) — Substitute the following for the existing clause:

13.1.1 BIS Certification Mark

The pipette may also be marked with the Standard Mark.

13.1.1.1 The use of the Standard Mark is governed by the provisions of *Bureau* of *Indian Standards Act*, 1986 and the Rules and Regulations made thereunder. The details of conditions under which the licence for the use of Standard Mark may be granted to manufacturers or producers may be obtained from the Bureau of Indian Standards.'

(CHD 10)

Indian Standard

SPECIFICATION FOR GRADUATED PIPETTES

PART 3 PIPETTES FOR WHICH A WAITING TIME OF 15s IS SPECIFIED

(First Revision)

Laboratoryware and Related Apparatus Sectional Committee, CDC 33

Chairman

Representing

DR M. SANKAR DAS

Bhabha Atomic Research Centre, Trombay, Bombay

Members

SHRI S. V. GULAVANE (Alternate to Dr M. Sankar Das)

DR G. S. BAJWA

Defence Science Laboratory, Ministry of Defence (R&D), New Delhi

DR N. K. BANERJEE

Indian Agricultural Research Institute (ICAR). New Delhi

DR K. S. RAM (Alternate)

SHRI N. G. BASAK

Directorate General of Technical Development. New Delhi

SHRI I. K. KAPOOR (Alternate)

SHRI M. L. BHAMBANI

National Test House, Calcutta

DR B. B. Pal (Alternate)

SHRI S. K. CHAKRABARTI

Development Commissioner. Small Scale Industries, New Delhi

SHRI S. R. MUNDARGI (Alternate)

LT-COL M. CHAKRABORTY

Directorate-General Armed Forces Medical Services, Ministry of Defence (DGAFMS). New Delhi

DR A. V. DIVEKAR

Bio-Science Corporation, Bombay Swastik Thermometer Co, Delhi

SHRI R. K. GADODIA SHRI D. S. GADODIA (Alternate)

SHRI VEDA PRAKASH GUPTA

Hicks Thermometers (India) Ltd. Aligarh

SHRI S. C. KAPILA (Alternate.)

DR R. C. KHERA SHRI D. C. KHERA (Alternate)

Pushpak Enterprises, Bombay

(Continued on page 2)

C Copyright 1986

BUREAU OF INDIAN STANDARDS

This publication is protected under the Indian Copyright Act (XIV of 1957) and reproduction in whole or in part by any means except with written permission of the publisher shall be deemed to be an infringement of copyright under the said Act.

IS: 4162 (Part 3) - 1985

(Continued from page 1)

Members

Representing

SHRI MOHINDER NATH

National Physical Laboratory (CSIR), New

Delhi

SHRI B. G. MATHUR (Alternate)

Shri P. R. Rao Borosil Glass Works Ltd, Bombay

SHRI U. T. TEMBE (Alternate)

Modern Instruments, Bangalore Shri H. N. Ravi

DR F. L. SALDHANA Haffkine Institute, Bombay

SHRI A. B. CHUNODKAR (Alternate) SHRI BHANU K. SAMPAT

The Scientific Indian Glass Co Ltd. Calcutta

SHRI D. R. SEN (Alternate)

Shri S. K. Sehgal Tarsons Products Ltd, Calcutta

SHRI J. C. SHANDILYA Top Syringe Manufacturing Co. Bombay

SHRI J. K. WAD (Alternate) SHRI G. S. SHUKLA Ministry of Defence (DGI), New Delhi

SHRI K. K. AGNIHOTRI (Alternate)

Central Glass & Ceramic Research Institute DR B. C. SINHA (CSIR), Calcutta

SHRI R. SEN (Alternate)

India Meteorological Department, New Delhi SHRI G. P. SRIVASTAVA

SHRI S. GOPI NATH (Alternate)

SHRI SATISH CHANDER. Director General, ISI (Ex-officio Member)

Director (Chem)

Secretary

SHRI M. M. MALHOTRA Deputy Director (Chem), ISI

Volumetric Glassware Subcommittee, CDC 33: 1

Convener

SHRI P. R. RAO

Borosil Glass Works Ltd, Bombay

Members

SHRI U. T. TEMBE (Alternate to

Shri P. R. Rao)

SHRI S. V. GULAVANE

SHRI K. M. MANDAL

SHRI R. K. GUPTA (Alternate)

SHRI MOHINDER NATH

SHRI BHANU K. SAMPAT

SHRI D. R. SEN (Alternate) SHRI J. C. SHANDILYA

SHRI J. K. WAD (Alternate)

DR B. C. SINHA

SHRI S. VAIDYANATHAN

Bhabha Atomic Research Centre, Bombay

Ministry of Defence (DGI)

National Physical Laboratory (CSIR), New

Delhi The Scientific Indian Glass Co Ltd. Calcutta

Top Syringe Manufacturing Co. Bombay

Central Glass & Ceramic Research Institute

(CSIR), Calcutta
Vensil Glass Works Pvt Ltd, Bangalore

Indian Standard

SPECIFICATION FOR GRADUATED PIPETTES

PART 3 PIPETTES FOR WHICH A WAITING TIME OF 15s IS SPECIFIED

(First Revision)

O. FOREWORD

- 0.1 This Indian Standard (Part 3) (First Revision) was adopted by the Indian Standards Institution on 30 October 1985, after the draft finalized by the Laboratoryware and Related Apparatus Sectional Committee had been approved by the Chemical Division Council.
- 0.2 This standard was originally published as IS: 4162-1967*. Subsequently International Organization for Standardization brought out the following Standards:

ISO: 835/1-1981 Laboratory glassware — Graduated pipettes:
Part 1 General requirements

ISO: 835/2-1981 Laboratory glassware — Graduated pipettes:
Part 2 Pipettes for which no waiting time is specified

ISO: 835/3-1981 Laboratory glassware — Graduated pipettes:
Part 3. Pipettes for which a waiting time of 15s is specified

ISO: 835/4-1981 Laboratory glassware — Graduated pipettes:
Part 4 Blow-out pipettes.

The Committee responsible for the preparation of this standard decided to revise IS: 4162-1967* to align this standard with the above ISO standards. Accordingly this standard is now being revised in four parts: Part 1 General requirements; Part 2 Pipettes for which no waiting time is specified; Part 3 Pipettes for which a waiting time of 15s is specified; and Part 4 Blow-out pipettes.

- 0.3 In this revision the following have been included:
 - a) Method of setting of meniscus,
 - b) Procedure for measurement of capacity,

^{*}Specification for graduated pipettes.

IS: 4162 (Part 3) - 1985

- c) 0.5 ml capacity pipettes with Class A accuracy for type 4 only,
- d) Pipettes for which a waiting time of 15s has been specified,
- e) Pipettes known as blow-out pipettes, and
- f) Table for dimensions has been modified to bring in line with ISO Standards.
- 0.4 This standard contains 13.2 which calls for agreement between the purchaser and the supplier.
- 0.5 For the purpose of deciding whether a particular requirement of this standard is complied with, the final values, observed or calculated, expressing the result of a test or analysis, shall be rounded off in accordance with IS:2-1960*. The number of significant places retained in the rounded off values should be the same as that of the specified values in this standard.

1. SCOPE

- 1.1 This standard (Part 3) specifies the requirements for graduated pipettes with comparatively large jets, for which waiting time of 15s is specified in order to compensate for the possible effect of shorter delivery time, and which are suitable for general laboratory purposes. They are adjusted to Class A accuracy.
 - 1.1.1 The requirements specified are in conformity with IS: 8729-1977†.
 - 1.1.2 The requirements specified in Part 1 of this standard shall apply.

2. TERMINOLOGY

2.1 For the purpose of this standard definitions given in Part 1 of the standard, shall apply.

3. TYPES OF PIPETTES

- 3.1 Graduated pipettes for which a waiting time of 15s is specified shall be of one type as indicated below:
 - Type 4 Graduated pipettes adjusted for delivery of a liquid from zero line at the top to any graduation line, nominal capacity obtained by delivery down to the jet, 15s waiting time specified for accuracy Class A only.

^{*}Rules for rounding off numerical values (revised).

[†]Principles of construction and adjustment of volumetric glassware.

4. ACCURACY

4.1 The graduated pipettes shall have A Class of accuracy only and shall conform to the requirement of 4, Part 1 of this standard.

5. NOMINAL CAPACITY

5.1 The graduated pipettes shall be 0.5 ml, 1 ml, 2 ml, 5 ml, 10 ml and 25 ml nominal capacities respectively.

5.2 Procedure for Measurement of Capacity

5.2.1 The procedure given in 5.2, Part 1 of this standard shall be followed.

6. LIMITS OF VOLUMETRIC ERROR

6.1 Errors in the delivered Volume of Class A graduated pipettes shall not exceed the limits shown in Table 1, Part 1 of this standard.

7. CONSTRUCTION

7.1 Material

- 7.1.1 Graduated pipettes shall be constructed from glass of suitable chemical and thermal properties, shall be as far as possible free from visible defects and shall be reasonably free from internal stress. When graded according to the method prescribed in IS:2303-1963* the pipettes shall conform to Type 1 of the glass.
- 7.2 Dimensions Graduated pipettes shall comply with the dimensional requirements given in Table 2, Part 1 of this standard.

7.3 Finish, Top of Pipettes and Delivery Jet

7.3.1 Graduated pipettes shall comply with the requirements given in 7.3, 7.3.1 and 7.3.2, Part 1 of this standard.

8. GRADUATION AND FIGURING

- 8.1 Graduated pattern 1 or 2 of 8.3.1 and 8.3.2, Part 1 of this standard, shall be applied and typical figuring of graduation shall be as shown in Fig. 1.
- 8.2 Lowest graduation and lowest figured graduation lines shall be in accordance with Table 1.

9. PERMANENCY AND VISIBILITY OF GRADUATION LINES, FIGURES INSCRIPTIONS AND MARKINGS

9.1 Graduation pipettes shall comply with the requirements given in 9, Part 1 of this standard.

^{*}Method of grading glass of alkalinity.




FIG. 1 GRADUATION AND FIGURING OF CLASS 'A', PIPETTES, TYPE 4, PATTERN 2

TABLE 1 LOWER-END OF SCALE

(Clause 8.2)

| SL No. | Nominal Capacity | Lowest Figures GRADUATION LINE | Lowest Graduation Line |
|-----------|------------------|--------------------------------|------------------------------|
| (1) | (2) | (3) | (4) |
| | ml | ml | ml |
| i) | 0.5 | 0.5 | 0.45 |
| ii) | 1.0 | 0.9 | 0.9 |
| iii) | 2.0 | 1.8 | 1.8 |
| iv) | 5∙0 | 4.5 | 4.5 |
| v) | 10.0 | 8.0 | 8.5 |
| vi) | 25.0 | 23.0 | 23·2 |

10. SETTING OF THE MENISCUS

10.1 Setting of the meniscus shall be performed in accordance with 10, Part 1 of this standard.

11. DELIVERY TIME

11.1 When the delivery time is marked on the pipette, then the observed delivery time and the marked delivery time shall both be within the limits given in Table 2 and shall not differ by more than 2s.

| | TABLE 2 DELIVERY T | IMES | |
|----------|--------------------|------|--|
| NOMINAL | DELIVERY TIMES | | |
| CAPACITY | Min | Max | |
| mi | S | 8 | |
| 0.5 | 2 | 8 | |
| 1 | 2 | 8 | |
| 2 | 2 | 8 | |
| 5 | 5 | 11 | |
| 10 | 5 | 11 | |
| 25 | 9 | 15 | |

12. COLOUR CODING

12.1 If colour coding is used on the graduated pipettes, it shall comply with the requirements of IS: 11468-1985* and IS: 11469-1985*.

^{*}Specification for colour coding of pipettes.

[†]Method of assesseny chemical resistance of enamels used in colour coding.

IS: 4162 (Part 3) - 1985

13. MARKING AND PACKING

- 13.1 Each pipettes shall be marked permanently and legibly on surface with the following information:
 - a) Nominal capacity followed by 'ml';
 - b) Letter 27°C to indicate the reference temperature;
 - c) Letter 'A' to indicate the class of accuracy for which the pipette has been adjusted;
 - d) Waiting time of 15s shall be indicated on the graduated pipettes;
 - e) Delivery time;
 - f) Type of pipette;
 - g) Letter 'Ex' to indicate that the pipette has been adjusted for delivery; and
 - h) Maker's name or recognized trade-mark, if any.
- 13.1.1 The pipette may also be marked with the ISI Certification Mark.

Note — The use of the ISI Certification Mark is governed by the provisions of the Indian Standards Institution (Certification Marks) Act and the Rules and Regulations made thereunder. The ISI Mark on the products covered by an Indian Standard conveys the assurance that they have been produced to comply with the requirements of that standard under a well-defined system of inspection, testing and quality control which is devised and supervised by ISI and operated by the producer. ISI marked products and also continuously checked by ISI for conformity to that standard under as a further safeguard. Details of conditions under which a licence for the ISI Certification Mark may be granted to manufacturer or processors, may be obtained from the Indian Standards Institution.

13.2 The pipettes shall be packed as agreed to between the purchaser and the supplier.

14. SAMPLING

14.1 Representative samples of the graduated pipettes shall be drawn as prescribed in Appendix C of Part 1 of this standard.

BUREAU OF INDIAN STANDARDS

| Headquarters: | |
|---|---|
| Manak Bhavan, 9 Bahadur Shah Zafar Marg, NE | W DELHI 110002 |
| Telephones: 331 01 31, 331 13 75 Te | legrams: Manaksanstha Common to all Offices) |
| Regional Offices: | Telephone |
| Central Manak Bhavan, 9 Bahadur Shah Zafar NEW DELHI 110002 | Marg. (331 01 31 331 13 75 |
| *Eastern : 1/14 C. I. T. Scheme VII M, V. I. P. Maniktola, CALCUTTA 700054 | |
| Northern: SCO 445-446, Sector 35-C, CHANDIGARH 160036 | { 2 18 43 { 3 16 41 (41 24 42 |
| Southern: C. I. T. Campus, MADRAS 600113 | 41 25 19 41 29 16 |
| †Western: Manakalaya, E9 MIDC, Marol, Andho BOMBAY 400093 | eri (East), 6 32 92 95 |
| Branch Offices: | |
| 'Pushpak', Nurmohamed Shaikh Marg, Khanpur, AHMADABAD 380001 | (2 63 48 } 2 63 49 |
| ‡Peenya Industrial Area 1st Stage, Bangalore To BANGALORE 560058 | 38 49 55 (38 49 56) 38 49 56 |
| Gangotri Complex, 5th Floor, Bhadbhada Road, BHOPAL 462003 | • |
| Plot No. 82/83, Lewis Road, BHUBANESHWAR 53/5, Ward No. 29, R.G. Barua Road, 5th Byelar GUWAHATI 781003 | 751002 5 36 27 ne, 3 31 77 |
| 6-8-56C L. N. Gupta Marg (Nampally Station Re HYDERABAD 500001 | • |
| R14 Yudhister Marg, C Scheme, JAIPUR 30200 | 1 0 30 32 |
| 117/418 B Sarvodaya Nagar, KANPUR 208005 | 21 68 76 21 82 92 |
| Patliputra Industrial Estate, PATNA 800013 | 6 23 05 |
| T.C. No. 14/1421. University P.O., Palayam TRIVANDRUM 695035 | ∫6 21 04 \ 6 21 17 |
| Inspection Offices (With Sale Point): | |
| Pushpanjali, First Floor, 205-A West High Court Shankar Nagar Square, NAGPUR 440010 | |
| Institution of Engineers (India) Building, 1332 S PUNE 411005 | hivaji Nagar, 5 24 35 |

^{*}Sales Office in Calcutta is at 5 Chowringhee Approach, P. O. Princep 27 68 00 Street. Calcutta 700072

[†]Sales Office in Bombay is at Novelty Chambers, Grant Road, 89 65 28 Bombay 400007

[‡]Sales Office in Bangalore is at Unity Building, Nerasimharaja Square, 22 36 71 Bangalore 560002